

SAFETY DATA SHEET



Clinitek Atlas Calibration Kit

SDS #: 5018A

Section 1. Identification

Product identifier : Clinitek Atlas Calibration Kit
Product code : 5018A, 07261789, 10311123
Product type : Liquid.

Relevant identified uses of the substance or mixture and uses advised against

Not applicable.

Manufactured/supplied : Siemens Healthcare Diagnostics Inc.
511 Benedict Avenue
Tarrytown, NY 10591-5097 USA
1-877-229-3711
(800) 424-9300 (CHEMTREC) (24/365)

Section 2. Hazards identification

OSHA/HCS status	: Clinitek Atlas Calibrator 1	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910.1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
	Clinitek Atlas Calibrator 2	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910.1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
	Clinitek Atlas Calibrator 3	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910.1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
	Clinitek Atlas Calibrator 4	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910.1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
Classification of the substance or mixture	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not classified. Not classified. Not classified. Not classified.
Additional information	: Not available.	

Section 2. Hazards identification

Not available.

GHS label elements

Signal word	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	No signal word. No signal word. No signal word. No signal word.
Hazard statements	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards.
Precautionary statements		
Prevention	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not applicable. Not applicable. Not applicable. Not applicable.
Response	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not applicable. Not applicable. Not applicable. Not applicable.
Storage	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not applicable. Not applicable. Not applicable. Not applicable.
Disposal	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not applicable. Not applicable. Not applicable. Not applicable.
Supplemental label elements	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	None known. None known. None known. None known.
Hazards not otherwise classified	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	None known. None known. None known. None known.

Section 3. Composition/information on ingredients

Substance/mixture	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Mixture Mixture Mixture Mixture
--------------------------	--	--

Ingredient name	%	CAS number
Clinitek Atlas Calibrator 4 sodium azide	0.05	26628-22-8

Any concentration shown as a range is to protect confidentiality or is due to batch variation.

There are no additional ingredients present which, within the current knowledge of the supplier and in the concentrations applicable, are classified as hazardous to health or the environment and hence require reporting in this section.

Occupational exposure limits, if available, are listed in Section 8.

Section 4. First aid measures

Description of necessary first aid measures

Eye contact	: Clinitek Atlas Calibrator 1	Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.
	Clinitek Atlas Calibrator 2	Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.
	Clinitek Atlas Calibrator 3	Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.
	Clinitek Atlas Calibrator 4	Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.
Inhalation	: Clinitek Atlas Calibrator 1	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 2	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur. In case of inhalation of decomposition products in a fire, symptoms may be delayed. The exposed person may need to be kept under medical surveillance for 48 hours.
	Clinitek Atlas Calibrator 3	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur. In case of inhalation of decomposition products in a fire, symptoms may be delayed. The exposed person may need to be kept under medical surveillance for 48 hours.
	Clinitek Atlas Calibrator 4	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
Skin contact	: Clinitek Atlas Calibrator 1	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 2	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 3	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 4	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.

Section 4. First aid measures

Ingestion	: Clinitek Atlas Calibrator 1	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 2	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 3	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
	Clinitek Atlas Calibrator 4	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.

Most important symptoms/effects, acute and delayed

Potential acute health effects

Eye contact	: Clinitek Atlas Calibrator 1	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 2	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 3	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 4	No known significant effects or critical hazards.
Inhalation	: Clinitek Atlas Calibrator 1	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 2	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 3	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 4	No known significant effects or critical hazards.
Skin contact	: Clinitek Atlas Calibrator 1	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 2	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 3	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 4	No known significant effects or critical hazards.

Section 4. First aid measures

Ingestion	: Clinitek Atlas Calibrator 1	hazards.
	Clinitek Atlas Calibrator 2	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 3	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 4	No known significant effects or critical hazards.

Over-exposure signs/symptoms

Eye contact	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.
Inhalation	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.
Skin contact	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.
Ingestion	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.

Indication of immediate medical attention and special treatment needed, if necessary

Notes to physician	: In case of inhalation of decomposition products in a fire, symptoms may be delayed. The exposed person may need to be kept under medical surveillance for 48 hours.
Specific treatments	: No specific treatment.
Protection of first-aiders	: No action shall be taken involving any personal risk or without suitable training.

See toxicological information (Section 11)

Section 5. Fire-fighting measures

Extinguishing media

Suitable extinguishing media	: In case of fire, use water spray (fog), foam or dry chemical.
Unsuitable extinguishing media	: None known.

Specific hazards arising from the chemical : In a fire or if heated, a pressure increase will occur and the container may burst.

Hazardous thermal decomposition products	: Decomposition products may include the following materials: carbon dioxide carbon monoxide nitrogen oxides
---	---

Special protective actions for fire-fighters : Promptly isolate the scene by removing all persons from the vicinity of the incident if there is a fire. No action shall be taken involving any personal risk or without suitable training.

Special protective equipment for fire-fighters : Fire-fighters should wear appropriate protective equipment and self-contained breathing apparatus (SCBA) with a full face-piece operated in positive pressure mode.

Section 6. Accidental release measures

Personal precautions, protective equipment and emergency procedures

- For non-emergency personnel** : No action shall be taken involving any personal risk or without suitable training. Evacuate surrounding areas. Keep unnecessary and unprotected personnel from entering. Do not touch or walk through spilled material. Put on appropriate personal protective equipment.
- For emergency responders** : If specialized clothing is required to deal with the spillage, take note of any information in Section 8 on suitable and unsuitable materials. See also the information in "For non-emergency personnel".
- Environmental precautions** : Avoid dispersal of spilled material and runoff and contact with soil, waterways, drains and sewers. Inform the relevant authorities if the product has caused environmental pollution (sewers, waterways, soil or air).

Methods and materials for containment and cleaning up

- Small spill** : Stop leak if without risk. Move containers from spill area. Dilute with water and mop up if water-soluble. Alternatively, or if water-insoluble, absorb with an inert dry material and place in an appropriate waste disposal container. Dispose of via a licensed waste disposal contractor.
- Large spill** : Stop leak if without risk. Move containers from spill area. Prevent entry into sewers, water courses, basements or confined areas. Wash spillages into an effluent treatment plant or proceed as follows. Contain and collect spillage with non-combustible, absorbent material e.g. sand, earth, vermiculite or diatomaceous earth and place in container for disposal according to local regulations (see Section 13). Dispose of via a licensed waste disposal contractor. Note: see Section 1 for emergency contact information and Section 13 for waste disposal.

Section 7. Handling and storage

Precautions for safe handling

- Protective measures** : Put on appropriate personal protective equipment (see Section 8).
- Advice on general occupational hygiene** : Eating, drinking and smoking should be prohibited in areas where this material is handled, stored and processed. Workers should wash hands and face before eating, drinking and smoking. Remove contaminated clothing and protective equipment before entering eating areas. See also Section 8 for additional information on hygiene measures.
- Conditions for safe storage, including any incompatibilities** : Store in accordance with local regulations. Store in original container protected from direct sunlight in a dry, cool and well-ventilated area, away from incompatible materials (see Section 10) and food and drink. Keep container tightly closed and sealed until ready for use. Containers that have been opened must be carefully resealed and kept upright to prevent leakage. Do not store in unlabeled containers. Use appropriate containment to avoid environmental contamination.

Section 8. Exposure controls/personal protection

Control parameters

Occupational exposure limits

Ingredient name	Exposure limits
Clinitek Atlas Calibrator 4 sodium azide	ACGIH TLV (United States, 3/2015). Notes: as hydrazoic acid vapor C: 0.11 ppm, (as Hydrazoic acid vapor) Form: as Hydrazoic acid vapor ACGIH TLV (United States, 3/2015). C: 0.29 mg/m ³ , (as Sodium azide) Form: as Sodium azide NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: NAN3 CEIL: 0.3 mg/m ³ , (NAN3) NIOSH REL (United States, 10/2013).

Section 8. Exposure controls/personal protection

Absorbed through skin. Notes: as HN3
 CEIL: 0.1 ppm, (as HN3)
OSHA PEL 1989 (United States, 3/1989).
Absorbed through skin. Notes: as HN3
 CEIL: 0.1 ppm, (as HN3)
OSHA PEL 1989 (United States, 3/1989).
Absorbed through skin. Notes: as NaN3
 CEIL: 0.3 mg/m³, (as NaN3)

- Appropriate engineering controls** : Good general ventilation should be sufficient to control worker exposure to airborne contaminants.
- Environmental exposure controls** : Emissions from ventilation or work process equipment should be checked to ensure they comply with the requirements of environmental protection legislation. In some cases, fume scrubbers, filters or engineering modifications to the process equipment will be necessary to reduce emissions to acceptable levels.

Individual protection measures

- Hygiene measures** : Wash hands, forearms and face thoroughly after handling chemical products, before eating, smoking and using the lavatory and at the end of the working period. Appropriate techniques should be used to remove potentially contaminated clothing. Wash contaminated clothing before reusing. Ensure that eyewash stations and safety showers are close to the workstation location.
- Eye/face protection** : Safety eyewear complying with an approved standard should be used when a risk assessment indicates this is necessary to avoid exposure to liquid splashes, mists, gases or dusts. If contact is possible, the following protection should be worn, unless the assessment indicates a higher degree of protection: safety glasses with side-shields.
- Skin protection**
- Hand protection** : Chemical-resistant, impervious gloves complying with an approved standard should be worn at all times when handling chemical products if a risk assessment indicates this is necessary.
- Body protection** : Personal protective equipment for the body should be selected based on the task being performed and the risks involved and should be approved by a specialist before handling this product.
- Other skin protection** : Appropriate footwear and any additional skin protection measures should be selected based on the task being performed and the risks involved and should be approved by a specialist before handling this product.
- Respiratory protection** : Based on the hazard and potential for exposure, select a respirator that meets the appropriate standard or certification. Respirators must be used according to a respiratory protection program to ensure proper fitting, training, and other important aspects of use.

Section 9. Physical and chemical properties

- Physical state** : Clinitek Atlas Calibrator 1 Liquid.
 Clinitek Atlas Calibrator 2 Liquid.
 Clinitek Atlas Calibrator 3 Liquid.
 Clinitek Atlas Calibrator 4 Liquid.
- Color** : Clinitek Atlas Calibrator 1 Colorless.
 Clinitek Atlas Calibrator 2 Colorless.
 Clinitek Atlas Calibrator 3 Colorless.
 Clinitek Atlas Calibrator 4 Colorless.
- Odor** : Clinitek Atlas Calibrator 1 Odorless.
 Clinitek Atlas Calibrator 2 Odorless.
 Clinitek Atlas Calibrator 3 Odorless.
 Clinitek Atlas Calibrator 4 Odorless.
- pH** : Clinitek Atlas Calibrator 1 6.6 to 7
 Clinitek Atlas Calibrator 2 5 to 5.3
 Clinitek Atlas Calibrator 3 6 to 6.3
 Clinitek Atlas Calibrator 4 Not available.

Section 9. Physical and chemical properties

Flash point	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Flammability (solid, gas)	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not relevant/applicable due to nature of the product. Not relevant/applicable due to nature of the product. Not relevant/applicable due to nature of the product. Not relevant/applicable due to nature of the product.
Relative density	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	1.006 1.046 1.02 1
Solubility in water	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not relevant/applicable due to nature of the product. Not relevant/applicable due to nature of the product. Not relevant/applicable due to nature of the product. Not relevant/applicable due to nature of the product.
Partition coefficient: n-octanol/water	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Auto-ignition temperature	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Viscosity	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
<u>Aerosol product</u>		
Type of aerosol	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not applicable. Not applicable. Not applicable. Not applicable.

Section 10. Stability and reactivity

Reactivity	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	No specific test data related to reactivity available for this product or its ingredients. No specific test data related to reactivity available for this product or its ingredients. No specific test data related to reactivity available for this product or its ingredients. No specific test data related to reactivity available for this product or its ingredients.
Chemical stability	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	The product is stable. The product is stable. The product is stable. The product is stable.
Possibility of hazardous reactions	: Under normal conditions of storage and use, hazardous reactions will not occur.	

Section 10. Stability and reactivity

Conditions to avoid : No specific data.

Incompatible materials : No specific data.

Hazardous decomposition products : Under normal conditions of storage and use, hazardous decomposition products should not be produced.

Section 11. Toxicological information

Information on toxicological effects

Acute toxicity

Product/ingredient name	Result	Species	Dose	Exposure
Clinitek Atlas Calibrator 4 sodium azide	LD50 Dermal	Rabbit	20 mg/kg	-
	LD50 Dermal	Rat	50 mg/kg	-
	LD50 Oral	Rat	27 mg/kg	-

Conclusion/Summary : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Irritation/Corrosion

Not available.

Conclusion/Summary

Skin : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Eyes : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Respiratory : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Sensitization

Not available.

Conclusion/Summary

Skin : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Respiratory : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Mutagenicity

Not available.

Conclusion/Summary : Clinitek Atlas Calibrator 1 Not available.
 Clinitek Atlas Calibrator 2 Not available.
 Clinitek Atlas Calibrator 3 Not available.
 Clinitek Atlas Calibrator 4 Not available.

Carcinogenicity

Not available.

Section 11. Toxicological information

Conclusion/Summary	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Reproductive toxicity		
Not available.		
Conclusion/Summary	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Teratogenicity		
Not available.		
Conclusion/Summary	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Specific target organ toxicity (single exposure)		
Not available.		
Specific target organ toxicity (repeated exposure)		
Not available.		
Aspiration hazard		
Not available.		
Information on the likely routes of exposure	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	Not available. Not available. Not available. Not available.
Potential acute health effects		
Eye contact	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards.
Inhalation	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards.
Skin contact	: Clinitek Atlas Calibrator 1 Clinitek Atlas Calibrator 2 Clinitek Atlas Calibrator 3 Clinitek Atlas Calibrator 4	No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards. No known significant effects or critical hazards.

Section 11. Toxicological information

Ingestion	: Clinitek Atlas Calibrator 1	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 2	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 3	No known significant effects or critical hazards.
	Clinitek Atlas Calibrator 4	No known significant effects or critical hazards.

Symptoms related to the physical, chemical and toxicological characteristics

Eye contact	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.
Inhalation	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.
Skin contact	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.
Ingestion	: Clinitek Atlas Calibrator 1	No specific data.
	Clinitek Atlas Calibrator 2	No specific data.
	Clinitek Atlas Calibrator 3	No specific data.
	Clinitek Atlas Calibrator 4	No specific data.

Delayed and immediate effects and also chronic effects from short and long term exposure

Short term exposure

Potential immediate effects	: Clinitek Atlas Calibrator 1	Not available.
	Clinitek Atlas Calibrator 2	Not available.
	Clinitek Atlas Calibrator 3	Not available.
	Clinitek Atlas Calibrator 4	Not available.
Potential delayed effects	: Clinitek Atlas Calibrator 1	Not available.
	Clinitek Atlas Calibrator 2	Not available.
	Clinitek Atlas Calibrator 3	Not available.
	Clinitek Atlas Calibrator 4	Not available.

Long term exposure

Potential immediate effects	: Clinitek Atlas Calibrator 1	Not available.
	Clinitek Atlas Calibrator 2	Not available.
	Clinitek Atlas Calibrator 3	Not available.
	Clinitek Atlas Calibrator 4	Not available.
Potential delayed effects	: Clinitek Atlas Calibrator 1	Not available.
	Clinitek Atlas Calibrator 2	Not available.
	Clinitek Atlas Calibrator 3	Not available.
	Clinitek Atlas Calibrator 4	Not available.

Potential chronic health effects

Not available.

Conclusion/Summary	: Not available.	Clinitek Atlas Calibrator 1
	Not available.	Clinitek Atlas Calibrator 2
	Not available.	Clinitek Atlas Calibrator 3
	Not available.	Clinitek Atlas Calibrator 4
General	: No known significant effects or critical hazards.	
Carcinogenicity	: No known significant effects or critical hazards.	
Mutagenicity	: No known significant effects or critical hazards.	
Teratogenicity	: No known significant effects or critical hazards.	
Developmental effects	: No known significant effects or critical hazards.	

Section 11. Toxicological information

Fertility effects : No known significant effects or critical hazards.

Numerical measures of toxicity

Acute toxicity estimates

Route	ATE value
Clinitek Atlas Calibrator 2 Oral	113213.6 mg/kg

Interactive effects : Clinitek Atlas Calibrator 1 Not available.
Clinitek Atlas Calibrator 2 Not available.
Clinitek Atlas Calibrator 3 Not available.
Clinitek Atlas Calibrator 4 Not available.

Other information : Clinitek Atlas Calibrator 1 Not available.
Clinitek Atlas Calibrator 2 Not available.
Clinitek Atlas Calibrator 3 Not available.
Clinitek Atlas Calibrator 4 Not available.

Section 12. Ecological information

Toxicity

Product/ingredient name	Result	Species	Exposure
Clinitek Atlas Calibrator 4 sodium azide	Acute EC50 0.348 mg/l Fresh water	Algae - Pseudokirchneriella subcapitata	96 hours
	Acute EC50 4.2 to 6.2 mg/l Fresh water	Daphnia - Daphnia pulex - Larvae	48 hours
	Acute LC50 9000 µg/l Fresh water	Crustaceans - Gammarus lacustris	48 hours
	Acute LC50 0.68 mg/l Fresh water	Fish - Lepomis macrochirus	96 hours
	Chronic NOEC 5600 µg/l Marine water	Algae - Macrocystis pyrifera	96 hours

Conclusion/Summary : Clinitek Atlas Calibrator 1 Not available.
Clinitek Atlas Calibrator 2 Not available.
Clinitek Atlas Calibrator 3 Not available.
Clinitek Atlas Calibrator 4 Not available.

Persistence and degradability

Conclusion/Summary : Clinitek Atlas Calibrator 1 Not available.
Clinitek Atlas Calibrator 2 Not available.
Clinitek Atlas Calibrator 3 Not available.
Clinitek Atlas Calibrator 4 Not available.

Bioaccumulative potential

Not available.

Mobility in soil

Soil/water partition coefficient (K_{oc}) : Clinitek Atlas Calibrator 1 Not available.
Clinitek Atlas Calibrator 2 Not available.
Clinitek Atlas Calibrator 3 Not available.
Clinitek Atlas Calibrator 4 Not available.

Mobility : Clinitek Atlas Calibrator 1 Not available.
Clinitek Atlas Calibrator 2 Not available.
Clinitek Atlas Calibrator 3 Not available.
Clinitek Atlas Calibrator 4 Not available.

Other adverse effects : No known significant effects or critical hazards.

Section 13. Disposal considerations

Disposal methods : The generation of waste should be avoided or minimized wherever possible. Disposal of this product, solutions and any by-products should at all times comply with the requirements of environmental protection and waste disposal legislation and any regional local authority requirements. Dispose of surplus and non-recyclable products via a licensed waste disposal contractor. Waste should not be disposed of untreated to the sewer unless fully compliant with the requirements of all authorities with jurisdiction. Waste packaging should be recycled. Incineration or landfill should only be considered when recycling is not feasible. This material and its container must be disposed of in a safe way. Empty containers or liners may retain some product residues. Avoid dispersal of spilled material and runoff and contact with soil, waterways, drains and sewers.

Section 14. Transport information

DOT Classification

UN number	Clinitek Atlas Calibrator 1	Not regulated.
	Clinitek Atlas Calibrator 2	Not regulated.
	Clinitek Atlas Calibrator 3	Not regulated.
	Clinitek Atlas Calibrator 4	Not regulated.
UN proper shipping name	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Transport hazard class(es)	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Packing group	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Environmental hazards	Clinitek Atlas Calibrator 1	No.
	Clinitek Atlas Calibrator 2	No.
	Clinitek Atlas Calibrator 3	No.
	Clinitek Atlas Calibrator 4	No.
Additional information	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

TDG Classification

UN number	Clinitek Atlas Calibrator 1	Not regulated.
	Clinitek Atlas Calibrator 2	Not regulated.
	Clinitek Atlas Calibrator 3	Not regulated.
	Clinitek Atlas Calibrator 4	Not regulated.
UN proper shipping name	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Transport hazard class(es)	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

Section 14. Transport information

Packing group	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Environmental hazards	Clinitek Atlas Calibrator 1	No.
	Clinitek Atlas Calibrator 2	No.
	Clinitek Atlas Calibrator 3	No.
	Clinitek Atlas Calibrator 4	No.
Additional information	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

ADR/RID

UN number	Clinitek Atlas Calibrator 1	Not regulated.
	Clinitek Atlas Calibrator 2	Not regulated.
	Clinitek Atlas Calibrator 3	Not regulated.
	Clinitek Atlas Calibrator 4	Not regulated.
UN proper shipping name	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Transport hazard class(es)	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

Packing group	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Environmental hazards	Clinitek Atlas Calibrator 1	No.
	Clinitek Atlas Calibrator 2	No.
	Clinitek Atlas Calibrator 3	No.
	Clinitek Atlas Calibrator 4	No.
Additional information	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

IMDG

UN number	Clinitek Atlas Calibrator 1	Not regulated.
	Clinitek Atlas Calibrator 2	Not regulated.
	Clinitek Atlas Calibrator 3	Not regulated.
	Clinitek Atlas Calibrator 4	Not regulated.
UN proper shipping name	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

Section 14. Transport information

Transport hazard class(es)

Packing group	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Environmental hazards	Clinitek Atlas Calibrator 1	No.
	Clinitek Atlas Calibrator 2	No.
	Clinitek Atlas Calibrator 3	No.
	Clinitek Atlas Calibrator 4	No.
Additional information	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

IATA

UN number	Clinitek Atlas Calibrator 1	Not regulated.
	Clinitek Atlas Calibrator 2	Not regulated.
	Clinitek Atlas Calibrator 3	Not regulated.
	Clinitek Atlas Calibrator 4	Not regulated.
UN proper shipping name	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Transport hazard class(es)	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

Packing group	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-
Environmental hazards	Clinitek Atlas Calibrator 1	No.
	Clinitek Atlas Calibrator 2	No.
	Clinitek Atlas Calibrator 3	No.
	Clinitek Atlas Calibrator 4	No.
Additional information	Clinitek Atlas Calibrator 1	-
	Clinitek Atlas Calibrator 2	-
	Clinitek Atlas Calibrator 3	-
	Clinitek Atlas Calibrator 4	-

Special precautions for user : Clinitek Atlas Calibrator 1

Clinitek Atlas Calibrator 2

Transport within user's premises: always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.

Transport within user's premises: always transport in closed containers that are upright and secure. Ensure that persons

Section 14. Transport information

transporting the product know what to do in the event of an accident or spillage.

Clinitek Atlas Calibrator 3

Transport within user's premises: always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.

Clinitek Atlas Calibrator 4

Transport within user's premises: always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.

Transport in bulk according to Annex II of MARPOL and the IBC Code

Proper shipping name :
Ship type :
Pollution category :

Section 15. Regulatory information

U.S. Federal regulations : TSCA 8(a) PAIR: Poly(oxy-1,2-ethanediyl), α -[(1,1,3,3-tetramethylbutyl)phenyl]- ω -hydroxy-TSCA 8(a) CDR Exempt/Partial exemption: Not determined
United States inventory (TSCA 8b): Not determined.

Clean Air Act Section 112 (b) Hazardous Air Pollutants (HAPs) : Not listed

Clean Air Act Section 602 Class I Substances : Not listed

Clean Air Act Section 602 Class II Substances : Not listed

DEA List I Chemicals (Precursor Chemicals) : Not listed

DEA List II Chemicals (Essential Chemicals) : Not listed

SARA 302/304

Composition/information on ingredients

Name	%	EHS	SARA 302 TPQ		SARA 304 RQ	
			(lbs)	(gallons)	(lbs)	(gallons)
Clinitek Atlas Calibrator 4 sodium azide	0.05	Yes.	500	-	1000	-

SARA 304 RQ : 8000000 lbs / 3632000 kg

SARA 311/312

Classification : Not applicable.

Composition/information on ingredients

Section 15. Regulatory information

Name	%	Fire hazard	Sudden release of pressure	Reactive	Immediate (acute) health hazard	Delayed (chronic) health hazard
Clinitek Atlas Calibrator 4 sodium azide	0.05	No.	No.	No.	Yes.	No.

State regulations

- Massachusetts** : None of the components are listed.
New York : None of the components are listed.
New Jersey : None of the components are listed.
Pennsylvania : None of the components are listed.

International regulations

- Chemical Weapons Convention List Schedule I Chemicals** : Clinitek Atlas Calibrator 1 Not listed
 Clinitek Atlas Calibrator 2 Not listed
 Clinitek Atlas Calibrator 3 Not listed
 Clinitek Atlas Calibrator 4 Not listed
- Chemical Weapons Convention List Schedule II Chemicals** : Clinitek Atlas Calibrator 1 Not listed
 Clinitek Atlas Calibrator 2 Not listed
 Clinitek Atlas Calibrator 3 Not listed
 Clinitek Atlas Calibrator 4 Not listed
- Chemical Weapons Convention List Schedule III Chemicals** : Clinitek Atlas Calibrator 1 Not listed
 Clinitek Atlas Calibrator 2 Not listed
 Clinitek Atlas Calibrator 3 Not listed
 Clinitek Atlas Calibrator 4 Not listed

Section 16. Other information

History

Date of issue/Date of revision : 5/12/2017

Version : 1.05

Key to abbreviations

- ATE = Acute Toxicity Estimate
 BCF = Bioconcentration Factor
 GHS = Globally Harmonized System of Classification and Labelling of Chemicals
 IATA = International Air Transport Association
 IBC = Intermediate Bulk Container
 IMDG = International Maritime Dangerous Goods
 LogPow = logarithm of the octanol/water partition coefficient
 MARPOL = International Convention for the Prevention of Pollution From Ships, 1973 as modified by the Protocol of 1978. ("Marpol" = marine pollution)
 UN = United Nations

▣ Indicates information that has changed from previously issued version.

Notice to reader

Allergen :