



# XN CHECK BF™

BODY FLUID CONTROL FOR SYSMEX XN-SERIES (XN-10, XN-20) ANALYZERS

Expiration Date:	<b>2019-08-18</b>	Quality Control Data Due Date #1:	8-Jul-19	Quality Control Data Due Date #2:	18-Aug-19
------------------	-------------------	--------------------------------------	----------	--------------------------------------	-----------

Lot Number:	<b>91471301</b>	<b>91471302</b>
-------------	-----------------	-----------------

Control:	<b>L1:Level 1</b>	<b>L2:Level 2</b>
----------	-------------------	-------------------

PARAMETERS	MEAN	EXPECTED RANGE	MEAN	EXPECTED RANGE
WBC-BF( $10^3/\mu\text{L}$ )	0.080	0.046 - 0.114	0.320	0.230 - 0.410
RBC-BF( $10^6/\mu\text{L}$ )	0.025	0.020 - 0.030	0.076	0.065 - 0.087
TC-BF#( $10^3/\mu\text{L}$ )	0.080	0.046 - 0.114	0.320	0.230 - 0.410
MN#( $10^3/\mu\text{L}$ )	0.018	0.010 - 0.026	0.075	0.050 - 0.101
PMN#( $10^3/\mu\text{L}$ )	0.062	0.029 - 0.095	0.245	0.167 - 0.323
MN%	23.0	14.7 - 31.3	23.3	18.4 - 28.2
PMN%	77.0	60.8 - 93.2	76.7	66.0 - 87.4

Please refer to the package insert for the handling of XN CHECK BF.